



## 8-2020 ADDENDUM 1

### SUPPLY AND DELIVERY OF LABORATORY GASES

#### **URGENT**

**PLEASE FORWARD THIS DOCUMENT TO  
WHOEVER IS IN POSSESSION OF THE BID  
OPPORTUNITY**

ISSUED: January 20, 2020  
BY: Brock Bradford  
TELEPHONE NO. 204 986-4752

**THIS ADDENDUM SHALL BE INCORPORATED  
INTO THE BID OPPORTUNITY AND SHALL  
FORM A PART OF THE CONTRACT  
DOCUMENTS**

Template Version: A20160708

---

**Please note the following and attached changes, corrections, additions, deletions, information and/or instructions in connection with the Bid Opportunity, and be governed accordingly. Failure to acknowledge receipt of this Addendum in Paragraph 8 of Form A: Bid may render your Bid non-responsive.**

---

#### **PART E – SPECIFICATIONS**

Revise E2.4 (b) to read:  
E2.4 (b) Minimum Carbon Dioxide purity = 99.8%;

#### **Questions and Answers**

Q1. Can you please confirm if a minimum purity of 99.999% is required for specification E2.4?

A1. The product is required to have a minimum purity of 99.8% for Item No. 3 - Carbon Dioxide Gas Syphon